

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082208.D
 Acq On : 20 May 2024 14:34
 Operator : JC\MD
 Sample : P2529-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-191D2-GW-20240514

Quant Time: May 21 03:47:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	306279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	142723	52.222	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.440%	
35) Dibromofluoromethane	8.165	113	100012	53.082	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.160%	
50) Toluene-d8	10.570	98	398816	55.211	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.420%	
62) 4-Bromofluorobenzene	12.847	95	131692	51.079	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 102.160%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.371	101	20922	10.287	ug/l	98
12) 1,1-Dichloroethene	4.341	96	6128	3.242	ug/l #	78
27) cis-1,2-Dichloroethene	7.477	96	3323	1.367	ug/l	95
30) Chloroform	7.965	83	2251	0.541	ug/l	85
38) Carbon Tetrachloride	8.359	117	3326	1.038	ug/l #	73
44) Trichloroethene	9.353	130	503263	235.119	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082208.D
Acq On : 20 May 2024 14:34
Operator : JC\MD
Sample : P2529-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-191D2-GW-20240514

Quant Time: May 21 03:47:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

