

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082209.D
 Acq On : 20 May 2024 14:58
 Operator : JC\MD
 Sample : P2529-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-191D1-GW-20240516

Quant Time: May 21 03:48:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	296406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	141092	53.155	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.300%			
35) Dibromofluoromethane	8.171	113	97805	53.640	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.280%			
50) Toluene-d8	10.571	98	382539	54.722	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.440%			
62) 4-Bromofluorobenzene	12.853	95	128565	51.527	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 103.060%			
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.371	101	8564	4.336	ug/l	93
12) 1,1-Dichloroethene	4.342	96	2107	1.148	ug/l #	77
16) Acetone	4.442	43	2563	2.354	ug/l	98
30) Chloroform	7.971	83	2171	0.537	ug/l	97
38) Carbon Tetrachloride	8.365	117	1229	0.396	ug/l #	92
44) Trichloroethene	9.353	130	136822	66.051	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082209.D
Acq On : 20 May 2024 14:58
Operator : JC\MD
Sample : P2529-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-191D1-GW-20240516

Quant Time: May 21 03:48:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

