

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082212.D
 Acq On : 20 May 2024 16:10
 Operator : JC\MD
 Sample : P2529-02DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-191D2-GW-20240514DL

Quant Time: May 21 03:49:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	291427	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	273114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115318	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	137117	52.607	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.220%	
35) Dibromofluoromethane	8.171	113	96558	53.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.720%	
50) Toluene-d8	10.571	98	375321	54.607	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.220%	
62) 4-Bromofluorobenzene	12.853	95	124200	50.628	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 101.260%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.377	101	3625	1.869	ug/l	94
12) 1,1-Dichloroethene	4.342	96	1104	0.612	ug/l	90
44) Trichloroethene	9.353	130	93337	45.828	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082212.D
Acq On : 20 May 2024 16:10
Operator : JC\MD
Sample : P2529-02DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-191D2-GW-20240514DL

Quant Time: May 21 03:49:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

