

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082216.D
 Acq On : 20 May 2024 17:47
 Operator : JC\MD
 Sample : P2513-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-15

Quant Time: May 21 03:51:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	292427	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	282100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	138741	53.088	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.180%	
35) Dibromofluoromethane	8.171	113	98241	54.612	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.220%	
50) Toluene-d8	10.570	98	379559	55.034	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.060%	
62) 4-Bromofluorobenzene	12.853	95	124866	50.725	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 101.460%	
Target Compounds						Qvalue
16) Acetone	4.430	43	25469	23.757	ug/l	93
20) Methylene Chloride	5.277	84	10198	4.715	ug/l #	81
43) Isopropyl Acetate	8.688	43	259688	43.406	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082216.D
Acq On : 20 May 2024 17:47
Operator : JC\MD
Sample : P2513-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-15

Quant Time: May 21 03:51:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

