

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082217.D
 Acq On : 20 May 2024 18:11
 Operator : JC\MD
 Sample : P2513-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-16

Quant Time: May 21 03:51:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166661	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	303230	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	120521	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	142765	54.250	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.500%	
35) Dibromofluoromethane	8.171	113	98813	52.973	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.940%	
50) Toluene-d8	10.565	98	391644	54.764	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.520%	
62) 4-Bromofluorobenzene	12.847	95	131491	51.514	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 103.020%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	20236	18.745	ug/l	92
20) Methylene Chloride	5.283	84	9997	4.590	ug/l #	77
43) Isopropyl Acetate	8.688	43	243589	39.265	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082217.D
Acq On : 20 May 2024 18:11
Operator : JC\MD
Sample : P2513-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-16

Quant Time: May 21 03:51:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

