

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082218.D
 Acq On : 20 May 2024 18:35
 Operator : JC\MD
 Sample : P2513-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-17

Quant Time: May 21 03:52:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	170057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	306203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296068	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122782	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	142190	52.952	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.900%	
35) Dibromofluoromethane	8.171	113	99148	52.637	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.280%	
50) Toluene-d8	10.571	98	398257	55.148	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.300%	
62) 4-Bromofluorobenzene	12.853	95	134021	51.995	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 103.980%	
Target Compounds						Qvalue
16) Acetone	4.436	43	20701	18.793	ug/l	89
20) Methylene Chloride	5.283	84	10986	4.943	ug/l #	83
43) Isopropyl Acetate	8.688	43	237145	37.855	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082218.D
Acq On : 20 May 2024 18:35
Operator : JC\MD
Sample : P2513-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-17

Quant Time: May 21 03:52:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

