

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082219.D
 Acq On : 20 May 2024 19:00
 Operator : JC\MD
 Sample : P2518-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-MH-MM

Quant Time: May 21 03:52:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	292330	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	136312	53.588	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.180%	
35) Dibromofluoromethane	8.171	113	96738	53.795	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.580%	
50) Toluene-d8	10.571	98	380595	55.203	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.400%	
62) 4-Bromofluorobenzene	12.853	95	125200	50.878	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 101.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	20611	19.752	ug/l	93
20) Methylene Chloride	5.289	84	10824	5.141	ug/l #	84
43) Isopropyl Acetate	8.688	43	289581	48.419	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082219.D
Acq On : 20 May 2024 19:00
Operator : JC\MD
Sample : P2518-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-MH-MM

Quant Time: May 21 03:52:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

