

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082220.D
 Acq On : 20 May 2024 19:24
 Operator : JC\MD
 Sample : P2518-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-MH-LL

Quant Time: May 21 03:53:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161454	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	290316	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115344	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	136493	53.539	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.080%		
35) Dibromofluoromethane	8.171	113	95051	53.223	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.440%		
50) Toluene-d8	10.570	98	378878	55.335	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 110.680%		
62) 4-Bromofluorobenzene	12.853	95	123517	50.542	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 101.080%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	22776	21.778	ug/l #	88
20) Methylene Chloride	5.283	84	13454	6.376	ug/l	86
43) Isopropyl Acetate	8.694	43	273596	46.063	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082220.D
Acq On : 20 May 2024 19:24
Operator : JC\MD
Sample : P2518-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-MH-LL

Quant Time: May 21 03:53:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

