

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
 Data File : VN082226.D
 Acq On : 20 May 2024 21:48
 Operator : JC\MD
 Sample : P2537-04 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 451

Quant Time: May 21 03:55:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	167968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304367	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	121553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	142244	53.631	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.260%	
35) Dibromofluoromethane	8.165	113	99140	52.950	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.900%	
50) Toluene-d8	10.570	98	383681	53.450	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.900%	
62) 4-Bromofluorobenzene	12.853	95	129913	50.705	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 101.420%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052024\
Data File : VN082226.D
Acq On : 20 May 2024 21:48
Operator : JC\MD
Sample : P2537-04 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
451

Quant Time: May 21 03:55:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

