

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
 Data File : VN086708.D
 Acq On : 20 May 2025 11:15
 Operator : JC\MD
 Sample : VN0520WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0520WBL01

Quant Time: May 21 01:41:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	186714	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	362119	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	350316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	148008	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	132562	51.380	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.760%
35) Dibromofluoromethane	8.165	113	114436	53.064	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.120%
50) Toluene-d8	10.565	98	473676	53.002	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.000%
62) 4-Bromofluorobenzene	12.847	95	165798	52.541	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
Data File : VN086708.D
Acq On : 20 May 2025 11:15
Operator : JC\MD
Sample : VN0520WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0520WBL01

Quant Time: May 21 01:41:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

