

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
 Data File : VN086711.D
 Acq On : 20 May 2025 12:27
 Operator : JC\MD
 Sample : PB168068TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB168068TB

Quant Time: May 21 01:43:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	163975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	323576	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316147	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	131784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	120727	53.281	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.560%		
35) Dibromofluoromethane	8.165	113	99532	51.650	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.300%		
50) Toluene-d8	10.565	98	423448	53.026	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.060%		
62) 4-Bromofluorobenzene	12.847	95	150478	53.366	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 106.740%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	23655	27.909	ug/l	97
18) Methyl Acetate	5.030	43	3265	1.376	ug/l #	77
20) Methylene Chloride	5.271	84	5427	2.426	ug/l #	85
43) Isopropyl Acetate	8.688	43	26288	4.567	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
Data File : VN086711.D
Acq On : 20 May 2025 12:27
Operator : JC\MD
Sample : PB168068TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB168068TB

Quant Time: May 21 01:43:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

