

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
 Data File : VN086713.D
 Acq On : 20 May 2025 13:15
 Operator : JC\MD
 Sample : Q2071-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 L1-WC-8

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 21 01:44:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173685	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	343623	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	334441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	144036	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	126544	52.726	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.460%			
35) Dibromofluoromethane	8.165	113	108101	52.824	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.640%			
50) Toluene-d8	10.565	98	446399	52.639	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.280%			
62) 4-Bromofluorobenzene	12.847	95	160456	53.585	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 107.180%			
Target Compounds						
16) Acetone	4.430	43	26321	29.318	ug/l	98
18) Methyl Acetate	5.024	43	2903	1.155	ug/l #	68
20) Methylene Chloride	5.271	84	9835	4.151	ug/l	92
43) Isopropyl Acetate	8.683	43	22336m	3.654	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
Data File : VN086713.D
Acq On : 20 May 2025 13:15
Operator : JC\MD
Sample : Q2071-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
L1-WC-8

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/21/2025
Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 21 01:44:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

