

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
 Data File : VN086717.D
 Acq On : 20 May 2025 14:51
 Operator : JC\MD
 Sample : Q2080-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-HRH-COMP-01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 21 01:46:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	150279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	299941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	128360	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112086	53.976	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.960%			
35) Dibromofluoromethane	8.165	113	92931	52.025	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.040%			
50) Toluene-d8	10.565	98	395364	53.410	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.820%			
62) 4-Bromofluorobenzene	12.847	95	141615	54.181	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 108.360%			
Target Compounds					Qvalue	
16) Acetone	4.424	43	82399	106.078	ug/l	94
18) Methyl Acetate	5.030	43	3367	1.549	ug/l #	71
20) Methylene Chloride	5.283	84	5671	2.767	ug/l	98
43) Isopropyl Acetate	8.688	43	18314m	3.432	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
Data File : VN086717.D
Acq On : 20 May 2025 14:51
Operator : JC\MD
Sample : Q2080-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-HRH-COMP-01

Quant Time: May 21 01:46:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/21/2025
Supervised By :Mahesh Dadoda 05/21/2025

