

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
 Data File : VN086718.D
 Acq On : 20 May 2025 15:15
 Operator : JC\MD
 Sample : Q2080-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-HRH-COMP-02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 21 01:46:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	168883	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	334715	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	315242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	134821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	123879	53.084	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.160%			
35) Dibromofluoromethane	8.165	113	103216	51.780	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.560%			
50) Toluene-d8	10.565	98	428611	51.886	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.780%			
62) 4-Bromofluorobenzene	12.847	95	147344	50.516	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.040%			
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	75110	86.042	ug/l	98
18) Methyl Acetate	5.024	43	3552	1.454	ug/l #	86
20) Methylene Chloride	5.283	84	5381	2.336	ug/l #	83
43) Isopropyl Acetate	8.683	43	18181m	3.053	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
Data File : VN086718.D
Acq On : 20 May 2025 15:15
Operator : JC\MD
Sample : Q2080-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-HRH-COMP-02

Quant Time: May 21 01:46:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/21/2025
Supervised By :Mahesh Dadoda 05/21/2025

