

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
 Data File : VN086719.D
 Acq On : 20 May 2025 15:39
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: May 21 01:47:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165892	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	327522	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	309617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	125758	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	122007	53.224	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.440%
35) Dibromofluoromethane	8.165	113	102604	52.603	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.200%
50) Toluene-d8	10.565	98	422523	52.272	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.540%
62) 4-Bromofluorobenzene	12.847	95	144135	50.501	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
Data File : VN086719.D
Acq On : 20 May 2025 15:39
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: May 21 01:47:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

