

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
 Data File : VN086706.D
 Acq On : 20 May 2025 10:10
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 21 01:39:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.540	0.599	-10.9	102	0.00
3 P	Chloromethane	0.817	0.781	4.4	102	0.00
4 C	Vinyl Chloride	0.733	0.732	0.1#	99	0.00
5 T	Bromomethane	0.423	0.380	10.2	97	0.01
6 T	Chloroethane	0.480	0.459	4.4	101	0.00
7 T	Trichlorofluoromethane	0.877	0.857	2.3	100	0.00
8 T	Diethyl Ether	0.362	0.351	3.0	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.525	0.540	-2.9	105	0.00
10 T	Methyl Iodide	0.653	0.586	10.3	92	0.00
11 T	Tert butyl alcohol	0.126	0.108	14.3	91	-0.01
12 CM	1,1-Dichloroethene	0.558	0.543	2.7#	102	0.00
13 T	Acrolein	0.135	0.119	11.9	91	0.00
14 T	Allyl chloride	0.844	0.831	1.5	103	0.00
15 T	Acrylonitrile	0.312	0.294	5.8	97	0.00
16 T	Acetone	0.258	0.230	10.9	97	0.00
17 T	Carbon Disulfide	1.544	1.541	0.2	102	0.00
18 T	Methyl Acetate	0.723	0.633	12.4	98	0.00
19 T	Methyl tert-butyl Ether	1.899	1.830	3.6	98	0.00
20 T	Methylene Chloride	0.682	0.622	8.8	101	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.579	1.9	103	0.00
22 T	Diisopropyl ether	1.904	1.917	-0.7	102	0.00
23 T	Vinyl Acetate	1.431	1.552	-8.5	109	0.00
24 P	1,1-Dichloroethane	1.118	1.094	2.1	101	0.00
25 T	2-Butanone	0.399	0.366	8.3	93	0.00
26 T	2,2-Dichloropropane	0.923	0.946	-2.5	109	0.00
27 T	cis-1,2-Dichloroethene	0.724	0.694	4.1	100	0.00
28 T	Bromochloromethane	0.508	0.511	-0.6	103	0.00
29 T	Tetrahydrofuran	0.261	0.244	6.5	94	0.00
30 C	Chloroform	1.119	1.067	4.6#	100	0.00
31 T	Cyclohexane	1.037	0.979	5.6	102	0.00
32 T	1,1,1-Trichloroethane	0.944	0.903	4.3	98	0.00
33 S	1,2-Dichloroethane-d4	0.691	0.684	1.0	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.298	0.296	0.7	100	0.00
36 T	1,1-Dichloropropene	0.439	0.437	0.5	101	0.00
37 T	Ethyl Acetate	0.444	0.410	7.7	95	0.00
38 T	Carbon Tetrachloride	0.433	0.428	1.2	100	0.00
39 T	Methylcyclohexane	0.475	0.497	-4.6	104	0.00
40 TM	Benzene	1.433	1.398	2.4	100	0.00
41 T	Methacrylonitrile	0.240	0.230	4.2	97	0.00
42 TM	1,2-Dichloroethane	0.446	0.430	3.6	99	0.00
43 T	Isopropyl Acetate	0.889	0.739	16.9	97	0.00
44 TM	Trichloroethene	0.346	0.346	0.0	101	0.00
45 C	1,2-Dichloropropane	0.340	0.343	-0.9#	103	0.00
46 T	Dibromomethane	0.230	0.226	1.7	100	0.00
47 T	Bromodichloromethane	0.473	0.469	0.8	100	0.00
48 T	Methyl methacrylate	0.345	0.331	4.1	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
 Data File : VN086706.D
 Acq On : 20 May 2025 10:10
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 21 01:39:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	93	0.00
50 S	Toluene-d8	1.234	1.258	-1.9	101	0.00
51 T	4-Methyl-2-Pentanone	0.449	0.418	6.9	93	0.00
52 CM	Toluene	0.891	0.879	1.3#	100	0.00
53 T	t-1,3-Dichloropropene	0.501	0.515	-2.8	101	0.00
54 T	cis-1,3-Dichloropropene	0.547	0.555	-1.5	102	0.00
55 T	1,1,2-Trichloroethane	0.332	0.314	5.4	99	0.00
56 T	Ethyl methacrylate	0.533	0.530	0.6	96	0.00
57 T	1,3-Dichloropropane	0.572	0.559	2.3	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.238	0.250	-5.0	97	0.00
59 T	2-Hexanone	0.328	0.306	6.7	93	0.00
60 T	Dibromochloromethane	0.348	0.346	0.6	100	0.00
61 T	1,2-Dibromoethane	0.334	0.322	3.6	97	0.00
62 S	4-Bromofluorobenzene	0.436	0.453	-3.9	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.383	0.387	-1.0	104	0.00
65 PM	Chlorobenzene	1.096	1.065	2.8	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.345	2.5	98	0.00
67 C	Ethyl Benzene	1.878	1.869	0.5#	99	0.00
68 T	m/p-Xylenes	0.711	0.726	-2.1	100	0.00
69 T	o-Xylene	0.707	0.700	1.0	96	0.00
70 T	Styrene	1.158	1.178	-1.7	97	0.00
71 P	Bromoform	0.252	0.248	1.6	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.862	3.520	8.9	99	0.00
74 T	N-amyl acetate	1.613	1.443	10.5	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.186	0.994	16.2	94	0.00
76 T	1,2,3-Trichloropropane	1.131	1.085	4.1	104	0.00
77 T	Bromobenzene	0.971	0.852	12.3	97	0.00
78 T	n-propylbenzene	4.432	4.236	4.4	100	0.00
79 T	2-Chlorotoluene	2.882	2.646	8.2	98	0.00
80 T	1,3,5-Trimethylbenzene	3.094	2.942	4.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.448	0.446	0.4	98	0.00
82 T	4-Chlorotoluene	2.855	2.696	5.6	101	0.00
83 T	tert-Butylbenzene	2.698	2.494	7.6	99	0.00
84 T	1,2,4-Trimethylbenzene	3.081	3.018	2.0	100	0.00
85 T	sec-Butylbenzene	3.618	3.542	2.1	102	0.00
86 T	p-Isopropyltoluene	2.965	3.006	-1.4	103	0.00
87 T	1,3-Dichlorobenzene	1.682	1.624	3.4	101	0.00
88 T	1,4-Dichlorobenzene	1.667	1.619	2.9	102	0.00
89 T	n-Butylbenzene	2.458	2.616	-6.4	106	0.00
90 T	Hexachloroethane	0.492	0.490	0.4	102	0.00
91 T	1,2-Dichlorobenzene	1.600	1.555	2.8	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.197	0.190	3.6	94	0.00
93 T	1,2,4-Trichlorobenzene	0.675	0.704	-4.3	102	0.00
94 T	Hexachlorobutadiene	0.280	0.283	-1.1	107	0.00
95 T	Naphthalene	2.345	2.256	3.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052025\
Data File : VN086706.D
Acq On : 20 May 2025 10:10
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 21 01:39:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.661	0.664	-0.5	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6