

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052119\
 Data File : VN055719.D
 Acq On : 21 May 2019 19:07
 Operator : JC/SP
 Sample : K2977-05 50X
 Misc : 6.12g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-W-5R(14-17)

Manual Integrations
 APPROVED

MMDadoda
 5/22/2019 4:10:53 PM

Quant Time: May 22 06:39:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 17 23:54:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	439582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	584638	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	510198	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	225538	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	180964	37.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.10%	
35) Dibromofluoromethane	7.59	113	176741	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
50) Toluene-d8	10.09	98	670106	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
62) 4-Bromofluorobenzene	12.40	95	230539	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.04	105	728433	54.101	ug/l	98
85) sec-Butylbenzene	13.17	105	117628	5.916	ug/l #	71
86) p-Isopropyltoluene	13.29	119	91510	4.944	ug/l	97
89) n-Butylbenzene	13.61	91	65969m	5.330	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052119\
Data File : VN055719.D
Acq On : 21 May 2019 19:07
Operator : JC/SP
Sample : K2977-05 50X
Misc : 6.12g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-W-5R(14-17)

Manual Integrations
APPROVED

MMDadoda
5/22/2019 4:10:53 PM

Quant Time: May 22 06:39:56 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051719W.M
Quant Title : SW846 8260
QLast Update : Fri May 17 23:54:12 2019
Response via : Initial Calibration

