

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052120\
 Data File : VN061525.D
 Acq On : 21 May 2020 13:07
 Operator : JC/MD
 Sample : PB129027TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129027TB

Quant Time: May 22 07:53:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	288606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	522744	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	523977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	226309	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	189480	43.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.98%	
35) Dibromofluoromethane	7.55	113	142002	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
50) Toluene-d8	10.06	98	648527	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
62) 4-Bromofluorobenzene	12.37	95	244851	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
Target Compounds						
20) Methylene Chloride	4.50	84	3669	1.036	ug/l #	81
43) Isopropyl Acetate	8.13	43	70467	9.114	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052120\
Data File : VN061525.D
Acq On : 21 May 2020 13:07
Operator : JC/MD
Sample : PB129027TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129027TB

Quant Time: May 22 07:53:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

