

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052120\
 Data File : VN061526.D
 Acq On : 21 May 2020 13:31
 Operator : JC/MD
 Sample : L2506-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-052020

Quant Time: May 22 07:56:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	284109	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	510222	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	495738	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	215004	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	185028	43.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.28%	
35) Dibromofluoromethane	7.55	113	142110	53.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.34%	
50) Toluene-d8	10.06	98	626834	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	12.37	95	232557	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
Target Compounds						
20) Methylene Chloride	4.50	84	133442	38.293	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052120\
Data File : VN061526.D
Acq On : 21 May 2020 13:31
Operator : JC/MD
Sample : L2506-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-052020

Quant Time: May 22 07:56:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

