

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052120\
 Data File : VN061528.D
 Acq On : 21 May 2020 14:17
 Operator : JC/MD
 Sample : PB129027ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129027ZHE#02

Quant Time: May 22 08:01:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	307347	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	566508	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	572577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	253524	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	200840	43.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.56%	
35) Dibromofluoromethane	7.55	113	147911	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
50) Toluene-d8	10.06	98	704501	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	12.37	95	270261	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
Target Compounds						
20) Methylene Chloride	4.51	84	11458	3.039	ug/l	93
43) Isopropyl Acetate	8.13	43	73129	8.728	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052120\
Data File : VN061528.D
Acq On : 21 May 2020 14:17
Operator : JC/MD
Sample : PB129027ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129027ZHE#02

Quant Time: May 22 08:01:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

