

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052120\
 Data File : VN061529.D
 Acq On : 21 May 2020 14:40
 Operator : JC/MD
 Sample : PB129027ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129027ZHE#03

Quant Time: May 22 08:03:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	293076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	539794	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	545960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	240189	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	195197	44.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.26%	
35) Dibromofluoromethane	7.55	113	142610	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	10.06	98	668454	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	12.37	95	259245	52.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.80%	
Target Compounds						
16) Acetone	3.78	43	13913	1.039	ug/l	94
20) Methylene Chloride	4.50	84	12572	3.497	ug/l	95
43) Isopropyl Acetate	8.13	43	73536	9.211	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052120\
Data File : VN061529.D
Acq On : 21 May 2020 14:40
Operator : JC/MD
Sample : PB129027ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129027ZHE#03

Quant Time: May 22 08:03:42 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

