

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052120\
 Data File : VN061530.D
 Acq On : 21 May 2020 15:04
 Operator : JC/MD
 Sample : PB129027ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129027ZHE#04

Quant Time: May 22 08:05:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	298127	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	546290	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	551975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	244271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	198318	44.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.14%	
35) Dibromofluoromethane	7.55	113	144050	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
50) Toluene-d8	10.06	98	683472	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
62) 4-Bromofluorobenzene	12.38	95	262626	52.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.92%	
Target Compounds						
16) Acetone	3.78	43	14619	1.469	ug/l	96
20) Methylene Chloride	4.51	84	12373	3.384	ug/l	91
43) Isopropyl Acetate	8.13	43	74705	9.246	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052120\
Data File : VN061530.D
Acq On : 21 May 2020 15:04
Operator : JC/MD
Sample : PB129027ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129027ZHE#04

Quant Time: May 22 08:05:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

