

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052120\
 Data File : VN061532.D
 Acq On : 21 May 2020 15:50
 Operator : JC/MD
 Sample : PB129027ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129027ZHE#06

Quant Time: May 22 08:09:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 18 13:09:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	301674	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	552694	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	557821	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	247954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	201509	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
35) Dibromofluoromethane	7.55	113	146081	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
50) Toluene-d8	10.06	98	691200	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
62) 4-Bromofluorobenzene	12.37	95	264235	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
Target Compounds						
16) Acetone	3.78	43	15366	1.990	ug/l	94
20) Methylene Chloride	4.51	84	12010	3.246	ug/l	91
43) Isopropyl Acetate	8.13	43	73519	8.994	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052120\
Data File : VN061532.D
Acq On : 21 May 2020 15:50
Operator : JC/MD
Sample : PB129027ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129027ZHE#06

Quant Time: May 22 08:09:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
Quant Title : SW846 8260
QLast Update : Mon May 18 13:09:55 2020
Response via : Initial Calibration

