


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN052120\
 Data File : VN061519.D
 Acq On : 21 May 2020 9:33
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051820W.M
 Title : SW846 8260
 Last Update : Mon May 18 13:09:55 2020



AutoFind: Scans 3296, 3297, 3298; Background Corrected with Scan 3284

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	6893	PASS
75	95	30	60	51.3	19408	PASS
95	95	100	100	100.0	37858	PASS
96	95	5	9	6.8	2568	PASS
173	174	0.00	2	1.4	382	PASS
174	95	50	100	71.3	26992	PASS
175	174	5	9	7.4	1995	PASS
176	174	95	101	101.0	27256	PASS
177	176	5	9	6.1	1650	PASS