


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052121\
 Data File : VN067156.D
 Acq On : 21 May 2021 13:11
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Title : SW846 8260
 Last Update : Fri May 21 17:24:08 2021



AutoFind: Scans 3960, 3961, 3962; Background Corrected with Scan 3944

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	4862	PASS
75	95	30	60	53.3	12801	PASS
95	95	100	100	100.0	24019	PASS
96	95	5	9	7.2	1725	PASS
173	174	0.00	2	0.4	69	PASS
174	95	50	100	69.1	16586	PASS
175	174	5	9	7.5	1244	PASS
176	174	95	101	96.1	15934	PASS
177	176	5	9	5.8	925	PASS