

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
 Data File : VN082241.D
 Acq On : 21 May 2024 15:29
 Operator : JC\MD
 Sample : P2532-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 22 01:27:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.229	168	157334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	282460	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	266636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	109735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	133733	53.830	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.660%
35) Dibromofluoromethane	8.171	113	92771	53.391	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.780%
50) Toluene-d8	10.570	98	365042	54.797	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.600%
62) 4-Bromofluorobenzene	12.847	95	121521	51.108	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	102.220%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
 Data File : VN082241.D
 Acq On : 21 May 2024 15:29
 Operator : JC\MD
 Sample : P2532-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 22 01:27:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

