

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
 Data File : VN082244.D
 Acq On : 21 May 2024 16:41
 Operator : JC\MD
 Sample : P2525-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-MH-AA

Quant Time: May 22 01:28:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	158754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	289471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	137055	54.674	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.340%		
35) Dibromofluoromethane	8.165	113	97496	54.752	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 109.500%		
50) Toluene-d8	10.571	98	381408	55.867	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 111.740%		
62) 4-Bromofluorobenzene	12.853	95	125512	51.508	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 103.020%		
Target Compounds						
						Qvalue
16) Acetone	4.442	43	29841	29.019	ug/l	90
20) Methylene Chloride	5.277	84	11097	5.349	ug/l #	82
43) Isopropyl Acetate	8.689	43	337686	57.020	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
Data File : VN082244.D
Acq On : 21 May 2024 16:41
Operator : JC\MD
Sample : P2525-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-MH-AA

Quant Time: May 22 01:28:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

