

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
 Data File : VN082248.D
 Acq On : 21 May 2024 18:18
 Operator : JC\MD
 Sample : P2539-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WCS-TPQ

Quant Time: May 22 01:30:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	160201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	291234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	285945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	135535	53.579	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.160%	
35) Dibromofluoromethane	8.171	113	96341	53.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.560%	
50) Toluene-d8	10.571	98	383897	55.891	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.780%	
62) 4-Bromofluorobenzene	12.853	95	125342	51.127	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 102.260%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	32614	31.429	ug/l	89
20) Methylene Chloride	5.277	84	10215	4.879	ug/l	88
43) Isopropyl Acetate	8.688	43	294314	49.395	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
Data File : VN082248.D
Acq On : 21 May 2024 18:18
Operator : JC\MD
Sample : P2539-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WCS-TPQ

Quant Time: May 22 01:30:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

