

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
 Data File : VN082258.D
 Acq On : 21 May 2024 23:08
 Operator : JC\MD
 Sample : VN0521MBL03
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0521MBL03

Quant Time: May 22 02:40:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	175418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	305159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	281416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115125	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	144624	52.213	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.420%
35) Dibromofluoromethane	8.171	113	102481	54.593	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.180%
50) Toluene-d8	10.571	98	388854	54.030	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.060%
62) 4-Bromofluorobenzene	12.847	95	127417	49.602	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	99.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
Data File : VN082258.D
Acq On : 21 May 2024 23:08
Operator : JC\MD
Sample : VN0521MBL03
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0521MBL03

Quant Time: May 22 02:40:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

