

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
 Data File : VN082266.D
 Acq On : 22 May 2024 02:21
 Operator : JC\MD
 Sample : P2494-06 40X
 Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 6

Quant Time: May 22 04:37:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	192934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	343247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	328265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	156615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	155326	50.985	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.980%	
35) Dibromofluoromethane	8.171	113	108106	51.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.400%	
50) Toluene-d8	10.570	98	435436	53.789	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.580%	
62) 4-Bromofluorobenzene	12.853	95	167181	57.860	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 115.720%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.606	83	8468	2.181	ug/l #	85
40) Benzene	8.606	78	26080	2.470	ug/l #	89
52) Toluene	10.635	92	45240	7.139	ug/l	96
67) Ethyl Benzene	11.965	91	30904	2.382	ug/l	93
68) m/p-Xylenes	12.070	106	65823	13.844	ug/l	89
69) o-Xylene	12.400	106	58447	12.691	ug/l	90
80) 1,3,5-Trimethylbenzene	13.176	105	100356	9.975	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	190787	18.536	ug/l	98
95) Naphthalene	15.641	128	2637795	295.309	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052124\
Data File : VN082266.D
Acq On : 22 May 2024 02:21
Operator : JC\MD
Sample : P2494-06 40X
Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
6

Quant Time: May 22 04:37:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

