

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
 Data File : VN086724.D
 Acq On : 21 May 2025 12:08
 Operator : JC\MD
 Sample : VN0521WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0521WBL01

Quant Time: May 22 01:20:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	174294	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	341228	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	334894	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	149145	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127718	53.030	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.060%
35) Dibromofluoromethane	8.159	113	108765	53.522	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.040%
50) Toluene-d8	10.565	98	443056	52.611	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.220%
62) 4-Bromofluorobenzene	12.847	95	161051	54.161	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
Data File : VN086724.D
Acq On : 21 May 2025 12:08
Operator : JC\MD
Sample : VN0521WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0521WBL01

Quant Time: May 22 01:20:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

