

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
 Data File : VN086730.D
 Acq On : 21 May 2025 14:41
 Operator : JC\MD
 Sample : PB168093TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB168093TB

Quant Time: May 22 01:23:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	206548	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	400057	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	399620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	175849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	139381	48.835	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.680%
35) Dibromofluoromethane	8.165	113	117206	49.194	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	10.565	98	529210	53.601	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.200%
62) 4-Bromofluorobenzene	12.847	95	198237	56.864	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	113.720%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	29917	28.022	ug/l	94
18) Methyl Acetate	5.030	43	3199	1.071	ug/l #	87
20) Methylene Chloride	5.283	84	6242	2.216	ug/l	94
43) Isopropyl Acetate	8.683	43	28824	4.050	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
Data File : VN086730.D
Acq On : 21 May 2025 14:41
Operator : JC\MD
Sample : PB168093TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB168093TB

Quant Time: May 22 01:23:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

