

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
 Data File : VN086738.D
 Acq On : 21 May 2025 17:54
 Operator : JC\MD
 Sample : Q2084-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-640-COMP-66

Quant Time: May 22 01:28:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	261016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	476055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	454248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	195313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	155981	43.247	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.500%
35) Dibromofluoromethane	8.165	113	140836	49.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	10.565	98	608052	51.754	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.500%
62) 4-Bromofluorobenzene	12.847	95	215193	51.873	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.740%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	20034	14.849	ug/l	98
18) Methyl Acetate	5.036	43	5352	1.417	ug/l	97
20) Methylene Chloride	5.277	84	6480	1.820	ug/l #	87
43) Isopropyl Acetate	8.683	43	29123	3.439	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
Data File : VN086738.D
Acq On : 21 May 2025 17:54
Operator : JC\MD
Sample : Q2084-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-640-COMP-66

Quant Time: May 22 01:28:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

