

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
 Data File : VN086739.D
 Acq On : 21 May 2025 18:18
 Operator : JC\MD
 Sample : Q2092-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-202

Quant Time: May 22 01:28:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	245107	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	456886	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	449469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	203767	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	153430	45.301	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.600%
35) Dibromofluoromethane	8.165	113	134850	49.560	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.120%
50) Toluene-d8	10.565	98	583437	51.743	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.480%
62) 4-Bromofluorobenzene	12.841	95	219923	55.238	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.480%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	20020	15.802	ug/l	96
18) Methyl Acetate	5.024	43	3552	1.002	ug/l #	84
20) Methylene Chloride	5.277	84	6560	1.962	ug/l #	87
43) Isopropyl Acetate	8.683	43	28780	3.541	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
Data File : VN086739.D
Acq On : 21 May 2025 18:18
Operator : JC\MD
Sample : Q2092-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-202

Quant Time: May 22 01:28:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

