

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
 Data File : VN086744.D
 Acq On : 21 May 2025 20:18
 Operator : JC\MD
 Sample : Q2032-01ME
 Misc : 6.89g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11ME

Quant Time: May 22 01:31:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	228680	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	425329	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	398917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	172654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	137758	43.595	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.200%
35) Dibromofluoromethane	8.159	113	124487	49.146	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	10.559	98	542032	51.637	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.280%
62) 4-Bromofluorobenzene	12.847	95	182379	49.206	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.420%
Target Compounds						
39) Methylcyclohexane	9.594	83	13149	3.252	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052125\
Data File : VN086744.D
Acq On : 21 May 2025 20:18
Operator : JC\MD
Sample : Q2032-01ME
Misc : 6.89g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11ME

Quant Time: May 22 01:31:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

