

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052219\
 Data File : VN055730.D
 Acq On : 22 May 2019 20:42
 Operator : JC/SP
 Sample : K3009-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 55-ST-14

Quant Time: May 23 06:33:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 23 05:54:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	55248	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	303694	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	244402	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	111228	30.80	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.67%
60) 4-Bromofluorobenzene	12.40	95	96744	25.45	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	84.83%
63) Toluene-d8	10.09	98	356666	31.93	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	106.43%

Target Compounds

					Ovalue
15) Acetone	3.82	58	4865	9.634 ug/l	98
22) cis-1,2-Dichloroethene	6.83	96	7108	1.943 ug/l	88
25) Chloroform	7.38	83	3221	0.566 ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052219\
Data File : VN055730.D
Acq On : 22 May 2019 20:42
Operator : JC/SP
Sample : K3009-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
55-ST-14

Quant Time: May 23 06:33:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 23 05:54:54 2019
Response via : Initial Calibration

