

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052219\
 Data File : VN055733.D
 Acq On : 22 May 2019 21:57
 Operator : JC/SP
 Sample : K2980-02 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-2

Quant Time: May 23 06:38:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 23 05:54:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	57973	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	304735	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	241916	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	114837	30.31	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.03%
60) 4-Bromofluorobenzene	12.40	95	92313	24.53	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	81.77%
63) Toluene-d8	10.09	98	360334	32.59	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	108.63%

Target Compounds

					Ovalue
25) Chloroform	7.37	83	130870	21.929	ug/l 99
41) Bromodichloromethane	9.40	83	19355	4.131	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052219\
Data File : VN055733.D
Acq On : 22 May 2019 21:57
Operator : JC/SP
Sample : K2980-02 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-2

Quant Time: May 23 06:38:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 23 05:54:54 2019
Response via : Initial Calibration

