

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052219\
 Data File : VN055734.D
 Acq On : 22 May 2019 22:22
 Operator : JC/SP
 Sample : K2980-03 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-3

Quant Time: May 23 06:40:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 23 05:54:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	57695	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	300549	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	241429	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	114471	30.35	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.17%
60) 4-Bromofluorobenzene	12.40	95	94570	25.18	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	83.93%
63) Toluene-d8	10.09	98	353415	32.03	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	106.77%

Target Compounds					Ovalue
25) Chloroform	7.37	83	136476	22.979 ug/l	100
41) Bromodichloromethane	9.40	83	20080	4.345 ug/l	98
53) Dibromochloromethane	10.90	129	1573	0.406 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052219\
Data File : VN055734.D
Acq On : 22 May 2019 22:22
Operator : JC/SP
Sample : K2980-03 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-3

Quant Time: May 23 06:40:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 23 05:54:54 2019
Response via : Initial Calibration

