

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052223\
 Data File : VN077779.D
 Acq On : 22 May 2023 10:56
 Operator : JC\MD
 Sample : VN0522MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0522MBL01

Quant Time: May 23 04:17:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	358195	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	633835	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	544658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	185482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	234744	43.799	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.600%
35) Dibromofluoromethane	8.177	113	207024	50.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.180%
50) Toluene-d8	10.576	98	756874	47.576	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.160%
62) 4-Bromofluorobenzene	12.859	95	232639	41.405	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.800%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052223\
Data File : VN077779.D
Acq On : 22 May 2023 10:56
Operator : JC\MD
Sample : VN0522MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0522MBL01

Quant Time: May 23 04:17:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

