

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052223\
 Data File : VN077792.D
 Acq On : 22 May 2023 16:17
 Operator : JC\MD
 Sample : 02880-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW8-MW01D1-20230518

Quant Time: May 23 04:20:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	389042	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	706225	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	599019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	194700	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	259935	44.654	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.300%
35) Dibromofluoromethane	8.177	113	225449	48.955	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.920%
50) Toluene-d8	10.577	98	853338	48.141	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.280%
62) 4-Bromofluorobenzene	12.859	95	259364	41.429	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.860%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052223\
Data File : VN077792.D
Acq On : 22 May 2023 16:17
Operator : JC\MD
Sample : 02880-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW8-MW01D1-20230518

Quant Time: May 23 04:20:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

