

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052223\
 Data File : VN077796.D
 Acq On : 22 May 2023 17:53
 Operator : JC\MD
 Sample : 02880-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT158S1-20230519

Quant Time: May 23 04:20:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	368315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	664514	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	574917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	186824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	243993	44.274	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.540%
35) Dibromofluoromethane	8.177	113	214337	49.464	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.920%
50) Toluene-d8	10.577	98	799638	47.943	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.880%
62) 4-Bromofluorobenzene	12.859	95	252733	42.904	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.800%
Target Compounds						
					Qvalue	
24) 1,1-Dichloroethane	6.577	63	15033	1.840	ug/l	98
44) Trichloroethene	9.353	130	1128	0.238	ug/l	42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052223\
Data File : VN077796.D
Acq On : 22 May 2023 17:53
Operator : JC\MD
Sample : 02880-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT158S1-20230519

Quant Time: May 23 04:20:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

