

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052223\
 Data File : VN077799.D
 Acq On : 22 May 2023 19:05
 Operator : JC\MD
 Sample : 02880-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT163S1-20230519

Quant Time: May 23 04:21:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	375536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	671062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	593624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	190636	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	250662	44.610	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.220%
35) Dibromofluoromethane	8.171	113	215251	49.190	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	10.576	98	820666	48.724	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.440%
62) 4-Bromofluorobenzene	12.859	95	256361	43.095	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.200%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.359	130	4776	0.999	ug/l	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052223\
Data File : VN077799.D
Acq On : 22 May 2023 19:05
Operator : JC\MD
Sample : 02880-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT163S1-20230519

Quant Time: May 23 04:21:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

