

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
 Data File : VN082290.D
 Acq On : 22 May 2024 13:53
 Operator : JC\MD
 Sample : PB161037ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161037ZHE#03

Quant Time: May 23 01:52:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	193684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	345798	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	328649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	138367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	157356	51.452	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.900%			
35) Dibromofluoromethane	8.165	113	110051	51.735	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.480%			
50) Toluene-d8	10.571	98	439307	53.866	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.740%			
62) 4-Bromofluorobenzene	12.847	95	154088	52.935	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 105.880%			
Target Compounds						
16) Acetone	4.436	43	22202	17.697	ug/l	96
20) Methylene Chloride	5.277	84	11131	4.397	ug/l #	81
43) Isopropyl Acetate	8.688	43	320237	45.265	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
Data File : VN082290.D
Acq On : 22 May 2024 13:53
Operator : JC\MD
Sample : PB161037ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161037ZHE#03

Quant Time: May 23 01:52:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

