

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
 Data File : VN082294.D
 Acq On : 22 May 2024 15:49
 Operator : JC\MD
 Sample : PB161037ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161037ZHE#07

Quant Time: May 23 01:53:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	185596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	330390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	316869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136647	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	151753	51.782	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.560%			
35) Dibromofluoromethane	8.171	113	107004	52.649	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.300%			
50) Toluene-d8	10.571	98	427750	54.895	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.800%			
62) 4-Bromofluorobenzene	12.853	95	152396	54.795	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 109.600%			
Target Compounds						
16) Acetone	4.436	43	33151	27.576	ug/l	93
20) Methylene Chloride	5.283	84	11005	4.537	ug/l #	77
43) Isopropyl Acetate	8.694	43	364367	53.905	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
Data File : VN082294.D
Acq On : 22 May 2024 15:49
Operator : JC\MD
Sample : PB161037ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161037ZHE#07

Quant Time: May 23 01:53:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

