

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
 Data File : VN082295.D
 Acq On : 22 May 2024 16:14
 Operator : JC\MD
 Sample : PB161037ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161037ZHE#08

Quant Time: May 23 01:53:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	190012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	337735	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	321680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	135227	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	154637	51.540	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.080%		
35) Dibromofluoromethane	8.171	113	107401	51.695	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.380%		
50) Toluene-d8	10.571	98	433019	54.363	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.720%		
62) 4-Bromofluorobenzene	12.853	95	151792	53.391	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 106.780%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	35545	28.880	ug/l	90
20) Methylene Chloride	5.277	84	10817	4.356	ug/l #	77
43) Isopropyl Acetate	8.694	43	358122	51.829	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
Data File : VN082295.D
Acq On : 22 May 2024 16:14
Operator : JC\MD
Sample : PB161037ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161037ZHE#08

Quant Time: May 23 01:53:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

