

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
 Data File : VN082296.D
 Acq On : 22 May 2024 16:38
 Operator : JC\MD
 Sample : PB161037ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161037ZHE#09

Quant Time: May 23 01:54:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	176112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	317122	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	299029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	121323	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	147474	53.032	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.060%	
35) Dibromofluoromethane	8.171	113	102057	52.316	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.640%	
50) Toluene-d8	10.571	98	405830	54.261	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.520%	
62) 4-Bromofluorobenzene	12.853	95	139062	52.093	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 104.180%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	34471	30.218	ug/l #	88
20) Methylene Chloride	5.283	84	11067	4.808	ug/l #	87
43) Isopropyl Acetate	8.694	43	339008	52.252	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
Data File : VN082296.D
Acq On : 22 May 2024 16:38
Operator : JC\MD
Sample : PB161037ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161037ZHE#09

Quant Time: May 23 01:54:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

