

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
 Data File : VN082300.D
 Acq On : 22 May 2024 18:15
 Operator : JC\MD
 Sample : PB161037ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161037ZHE#13

Quant Time: May 23 01:55:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	305576	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	287566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115906	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	142599	53.691	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.380%			
35) Dibromofluoromethane	8.171	113	97965	52.116	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.240%			
50) Toluene-d8	10.571	98	388999	53.976	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.960%			
62) 4-Bromofluorobenzene	12.853	95	135367	52.625	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 105.240%			
Target Compounds						
16) Acetone	4.436	43	35622	32.695	ug/l	94
20) Methylene Chloride	5.283	84	11009	5.008	ug/l #	81
43) Isopropyl Acetate	8.694	43	354474	56.700	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
Data File : VN082300.D
Acq On : 22 May 2024 18:15
Operator : JC\MD
Sample : PB161037ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161037ZHE#13

Quant Time: May 23 01:55:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

