

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
 Data File : VN082304.D
 Acq On : 22 May 2024 19:53
 Operator : JC\MD
 Sample : PB161037ZHE#17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161037ZHE#17

Quant Time: May 23 01:57:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	158758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	295980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	282296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	138453	55.230	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 110.460%		
35) Dibromofluoromethane	8.171	113	96300	52.891	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 105.780%		
50) Toluene-d8	10.571	98	382575	54.806	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 109.620%		
62) 4-Bromofluorobenzene	12.853	95	126139	50.627	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery	= 101.260%		
Target Compounds						Qvalue
16) Acetone	4.436	43	36542	35.535	ug/l	90
20) Methylene Chloride	5.277	84	12493	6.021	ug/l #	65
43) Isopropyl Acetate	8.694	43	332577	54.922	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
Data File : VN082304.D
Acq On : 22 May 2024 19:53
Operator : JC\MD
Sample : PB161037ZHE#17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161037ZHE#17

Quant Time: May 23 01:57:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

