

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
 Data File : VN082306.D
 Acq On : 22 May 2024 20:42
 Operator : JC\MD
 Sample : PB161037ZHE#19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161037ZHE#19

Quant Time: May 23 01:58:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	168827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	311787	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	304862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	131877	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	146470	54.944	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.880%	
35) Dibromofluoromethane	8.171	113	101826	53.090	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.180%	
50) Toluene-d8	10.571	98	407396	55.403	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.800%	
62) 4-Bromofluorobenzene	12.853	95	142757	54.392	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 108.780%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	35146	32.139	ug/l #	88
18) Methyl Acetate	5.030	43	4105	1.507	ug/l #	80
20) Methylene Chloride	5.283	84	11849	5.370	ug/l #	78
43) Isopropyl Acetate	8.688	43	318426	49.919	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052224\
Data File : VN082306.D
Acq On : 22 May 2024 20:42
Operator : JC\MD
Sample : PB161037ZHE#19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161037ZHE#19

Quant Time: May 23 01:58:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

