

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052319\
 Data File : VN055758.D
 Acq On : 23 May 2019 17:11
 Operator : JC/SP
 Sample : K3008-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-21

Quant Time: May 24 07:19:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	449736	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	622386	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	556419	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	218101	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	204906	44.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.24%	
35) Dibromofluoromethane	7.59	113	190055	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
50) Toluene-d8	10.09	98	738137	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	12.40	95	241896	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
Target Compounds						
16) Acetone	3.82	43	10694	5.238	ug/l	Qvalue 100
20) Methylene Chloride	4.54	84	105525	23.866	ug/l	99
43) Isopropyl Acetate	8.17	43	13258	1.505	ug/l #	89
95) Naphthalene	15.13	128	57256	5.675	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052319\
Data File : VN055758.D
Acq On : 23 May 2019 17:11
Operator : JC/SP
Sample : K3008-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-21

Quant Time: May 24 07:19:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

